

DIL 1.0 mL Aqua dest.

TECHNOZYM® Vitronectin Standard

| LOT | | Nominal value Sollwert µg/mL | calibrated against / kalibriert gegen |
|---------|------------|------------------------------------|--|
| RDR61B3 | 2023-08-31 | 405.3 | Internal reference plasma |

Standard Dilution:

| Tube | | Incubation buffer | Conc. % | Conc. µg/mL |
|-------------|---------------------|-------------------|------------|----------------|
| Predilution | 0.1 mL Standard | 0.9 mL | 10 | 40.53 |
| A | 0.01 mL Predilution | 0.99 mL | 0.1 | 0.405 |
| B | 0.5 mL from A | 0.5 mL | 0.05 | 0.203 |
| C | 0.5 mL from B | 0.5 mL | 0.025 | 0.101 |
| D | 0.5 mL from C | 0.5 mL | 0.0125 | 0.051 |
| E | - | 0.5 mL | 0 | 0 |

Qualitätskontrolle
quality control

1/1

