

**Ceveron TGA RC Low Kit**

**LOT B436C00**

**2024-07-31**

3013833- 020 20/11/2023

Ceveron 100 series barcode



Kontrolle control	Lot	Parameter	Einheit unit	Richtwert (Vertrauensbereich) expected value (confidence range)	Std.
Ceveron TGA Control High	1Y33C00	Peak thrombin	nM	261 ( 229 - 293 )	IS 19/188
Ceveron TGA Control Low	2734C00			181 ( 136 - 226 )	IS 19/188